
International Standards Certification Declaration

OCTOBER 28, 2016

Aqua Creek Products, LLC hereby declares that the Aqua Creek **Pro Pool Lift** (Model F-004PLB-AT1, F-004PLBNA-AT1, F-004PLB-R-AT1, F-004PLB-DD12-AT1, F-004PLB-23-AT1, F-004PLB-26-AT1), **Pro Pool XR Lift** (Models F-004PLB-XR, F-004PLB-XRNA, F-004PLB-XRNA-R) **Ranger Pool Lift** (model F-411RPL-AT1-EU), **Patriot Pool Lift** (Model F-12PPL-HD-AT1, F-12PPL-HD-AT2), **Portable Pro Pool Lift** (Model F-004PPPB-AT1), **Power EZ Pool Lift** (Model F-03PEZ, F-333PEZNA), **Super Power EZ Pool Lift** (Models F-190PHBL, F-194PHBL-NA), **ABLE Lift** (Model F-25ABLE), **Scout Pool Lift** (Model F-802-LIN-NA-EU), **Revolution Pool Lift** (Model F-700RL, F-702RLNA, F-702RLNA-DD, F-702RLNA-SPA), **Spa Lift** (Model F-005slb) **Pro Spa 40** (F-006PSLB), **Pro Spa 60** (F-007EXA), **Spa Lift Ultra** (F-005SLU), **Spa Lift Elite** (F-006SLE) and the **Titan Pool Lift** (Models F-600TL, F-600TLNA), all comply with the following Medical Equipment Directives:

31993L0042 Council Directive 93/42/EEC of 14 June 1993 concerning medical devices

EN ISO 14971:2009

Medical devices - Application of risk management to medical devices (ISO 14971:2007, Corrected version 2007-10-01)

EN ISO 10535:2006

Hoists for the transfer of disabled persons - Requirements and test methods (ISO 10535:2006) To conform with Directive 2007/47/EC as published

EN 60601-1-2:2001+A1:2006

Medical electrical equipment -- Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility - Requirements and tests
IEC 60601-1-2:2001

EN 60601-1:1990+A1:1993+A2:1995

Medical electrical equipment -- Part 1: General requirements for safety
IEC 60601-1:1988

Aqua Creek Products authorized representative:

Disability Pool Lifts, Australia
Attn: Clayton McLellan/Jacinta O'Connor
1/1927 Pittwater Rd
Bayview, NSW 2104
Australia
PH: 0411 055 109
jacinta@dplaustalia.com.au

Thank you



Brian Goettlich
CEO
Aqua Creek Products